

## DICHIARAZIONE DI CONFORMITA'



L'azienda Demarta Virginio S.a.S. dichiara, sotto la propria responsabilità, che i seguenti dispositivi medici marchiati CE:

Articolo	Codice
Comoda 600 P	01.04.103-600P
Comoda 600 A	01.04.103-600A
Comoda 4 x 200	01.04.103-4X200
Comoda 4 x 100	01.04.103-4X100
Comodina • 43 cm	01.04.100.43
Comodina • 46 cm	01.04.100.46
Comoda 3 in 1	01.04.31

sono classificati come classe di rischio I non sterile, in accordo alla regola 1 dell'Allegato IX della Direttiva 93/42/CEE.

Gli articoli soprariportati, inoltre:

- Rispettano le misure definite dai requisiti essenziali di salute e sicurezza della Direttiva 93/42/CEE (recepita in Italia con D.lgs 46/97), Allegato I e VII.
- Sono conformi a tutti gli standard europei ad essi applicabili.

Pollone, 20 Luglio 2018

DEMARTA VIRGINIO S.a.S.  
Dott. Virginio Belletti

A handwritten signature in black ink, appearing to read 'Virginio Belletti'.

## DECLARATION OF CONFORMITY



Demarta Virginio S.a.S. declares under his sole responsibility that the CE marked devices:

Article	Code
Commode Wheelchair 600 P	01.04.103-600P
Commode Wheelchair 600 A	01.04.103-600A
Commode Wheelchair 4 x 200	01.04.103-4X200
Commode Wheelchair 4 x 100	01.04.103-4X100
Wheeled Commode Chair • 43 cm	01.04.100.43
Wheeled Commode Chair • 46 cm	01.04.100.46
3 in 1 Commode Chair	01.04.31

have been classified as class I, according to annex IX MDD 93/42/EEC, rule 1.

The above-mentioned articles also comply with:

- The measures defined by the essential health and safety requirements of Directive 93/42/EEC (implemented in Italy with Decree 46/97) annex I and VII.
- All European standards applicable to them.

Pollone, 20 July 2018

DEMARTA VIRGINIO S.a.S.  
Dott. Virginio Belletti

A handwritten signature in black ink, appearing to read 'Virginio Belletti', written over a horizontal line.