



Product Service

# CERTIFICATE

No. B 07 02 13615 077

**Holder of Certificate:** **Semperit Technische Produkte Gesellschaft m.b.H.**

Modecenterstraße 22  
1031 Wien  
AUSTRIA

**Production Facility(ies):** 12175

**Certification Mark:**



**Product:** **Medical Gloves for single use**

**Model(s):** **Latex Surgical gloves steril "Sempermed Classic"**

Parameters:	Size	Article Number
	6,5	8260 54 629 and 8260 54 620
	7	8260 54 709 and 8260 54 700
	7,5	8260 54 729 and 8260 54 720
	8	8260 54 809 und 8260 54 800
	8,5	8260 54 829 and 8260 54 820

**Tested according to:** EN 455-1:2000  
EN 455-2:2000  
EN 455-3:1999

The product was tested on a voluntary basis and complies with the following essential requirements. The certification mark shown above can be affixed on the product. The certification mark must not be altered in any way. See also notes overleaf.

**Test report no.:** 028-70028880-100

**Date,** 2007-03-06

Page 1 of 1



## Declaration of conformity

This certificate is valid for all surgical gloves (classified IIa) described as follows and bearing the **CE 0123** mark

### **sempermed Classic 8260 54..0**

We hereby certify that the **CE 0123** marked products conform to the applicable provisions of the directive for medical devices 93/42 EEC.

This declaration is based on the conformity assessment according to annex II clause 3 of the same directive performed by the notified body TÜV Product Service with identification number 0123.  
Classification according rule 6, appendix IX.

Vienna,  
2005-20-10

  
Dkfm. G. Jirovsky  
Division Manager

  
Dr. M. Höchtl  
Quality Manager DM